

# HI-4453/FP5083/ FP4083/FP2083

## Isotropic Electric Field Probes User's Manual

### Declaration of Conformity

ETS-Lindgren, L.P.  
1301 Arrow Point Drive  
Cedar Park, TX 78613  
USA



Declares that the HOLADAY product described in this instruction manual is in compliance with: EN EMC Directive 86/336/EEC, EN50082-1, EN55011

ETS-Lindgren, L.P.  
(Formerly Holaday Industries, Inc.)



ETS-Lindgren L.P., reserves the right to make changes to any product described herein in order to improve function, design, or for any other reason. Nothing contained herein shall constitute ETS-Lindgren L.P. assuming any liability whatsoever arising out of the application or use of any product or circuit described herein. ETS-Lindgren L.P. does not convey any license under its patent rights or the rights of others.

© Copyright 2005 by ETS-Lindgren L.P. All Rights Reserved.

No part of this document may be copied by any means without written permission from ETS-Lindgren L.P.

**EMAIL & INTERNET ADDRESSES**

[support@ets-lindgren.com](mailto:support@ets-lindgren.com)

<http://www.ets-lindgren.com>

**USA**

1301 Arrow Point Dr., Cedar Park TX 78613 USA  
P.O. Box 80589, Austin, TX 78708-0589 USA

Tel 512.531.6400

Fax 512.531.6500

**FINLAND**

Mekaanikontie 1, 27510, Eura, Finland  
Tel 358.2.838.3300  
Fax 358.2.865.1233

**JAPAN**

4-2-6, Kohinata

Bunkyo-ku

Tokyo 112-0006 JAPAN

Phone + 81 3 3813 7100

Fax + 81 3 3813 8068

**CHINA**

1917-1918 Xue Zhixuan Building  
No 16 Xue Qing Road  
Haidian District  
Beijing Postcode: 100083 CHINA  
Phone + 86 010 82755304  
Fax + 86 010 82755307

---

Isotropic Electric Field Probes  
HI-4453/FP5083/FP4083/FP2083

**REVISION RECORD**

HI-4453/FP5083/FP4083/FP2083, MANUAL, Part #600064

| <b>Revision</b> | <b>Description</b>      | <b>Date</b>    |
|-----------------|-------------------------|----------------|
| A               | Initial Release         | October, 1993  |
| B               | Isotropicity Spec       | October, 1994  |
| C               | Changed Battery Charger | June, 1997     |
| D               | Added CE Label          | June, 1997     |
| E               | Changed Charger Specs   | August, 1999   |
| F               | Changed Area Code       | February, 2000 |
|                 | Revised                 | June, 2005     |

This page intentionally left blank.

## Table of Contents

|   |    |
|---|----|
| 1. Introduction.....  | 7  |
| Description .....   | 7  |
| Specifications .....  | 9  |
| 2. Getting Started.....   | 11 |
| Unpacking and Acceptance .....  | 11 |
| Probe .....   | 11 |
| Battery Charging .....  | 13 |
| 3. Operation.....   | 15 |
| System Theory .....   | 15 |
| Probe Operation.....  | 17 |
| Probe Power Supply.....   | 18 |
| Application Considerations.....   | 19 |
| 4. Maintenance.....   | 21 |
| Maintenance Recommendations.....  | 21 |
| Upgrade Policies .....  | 21 |
| Return Procedures .....   | 21 |
| Periodic/Preventive Maintenance.....  | 22 |
| Parts Information .....   | 23 |
| 5. Warranty Policy for Electromagnetic Isotropic<br>Field Monitoring Components ..... | 25 |
| Scope and Duration of Warranties .....  | 25 |
| Warranty Exclusions.....  | 26 |
| Buyer's Remedies .....  | 27 |
| 6. Appendix A: Error Codes .....  | 29 |
| Probe Error Output .....  | 29 |
| 7. Appendix B: HI-4453/FP5083/FP4083/FP2083<br>Operating Protocols .....              | 31 |
| Communication Protocol .....  | 31 |
| Information Transfer Protocol.....  | 31 |
| Commands .....  | 32 |
| 8. Appendix C: Series 491198-36 Battery Fast<br>Charger Manual.....                   | 35 |
| Introduction.....   | 45 |
| Power .....   | 47 |
| Output.....   | 47 |
| NiCd Battery: .....   | 47 |
| Environmental .....   | 47 |

---

Isotropic Electric Field Probes  
HI-4453/FP5083/FP4083/FP2083

|                                   |    |
|-----------------------------------|----|
| Power Cable .....                 | 47 |
| Power Cable Set Information ..... | 48 |
| Input Power Requirements .....    | 51 |
| Fuse Replacement .....            | 51 |
| Charging a Battery .....          | 52 |
| Charging Indicators .....         | 53 |

## 1. INTRODUCTION

### DESCRIPTION

The HI-4453/FP5083/FP4083/FP2083 Electric Field Probes embody the latest innovations in isotropic sensor design and low noise, miniaturized electronics. The HI-4453/FP5083/FP4083/FP2083 probes are fully intelligent sensor enabling fast and accurate EMF measurements with industry-leading performance specifications. Optical coupling to a variety of readout options makes this probe ideally suited for a wide range of field monitoring applications. The HI-4453/FP5083/FP4083/FP2083 probes are excellent tools for electric field mapping, RADHAZ measurements and EMC field monitoring.

The basic HI-4453/FP5083/FP4083/FP2083 probes are shipped complete with a 10 meter fiber extension cable, a carrying case, battery charger and connectors for extending the optic cable.

The HI-4453/FP5083/FP4083/FP2083 probe assembly consists of a spherical casing, containing the sensor, which is mounted on one end of a shaft; the other end of the shaft is attached to an extrusion that houses the electronics (Figure 1). The sensor and electronics housing operate, and are calibrated, as a unit.

The HI-4453/FP5083/FP4083/FP2083 measures the field strength in each of three axes. Frequency response of the HI-4453/FP5083/FP4083/FP2083 is 80 MHz to 40 GHz; dynamic range is 3 to 300 Volts per meter (V/m).

---

Isotropic Electric Field Probes  
HI-4453/FP5083/FP4083/FP2083



Figure 1: Model HI-4453/FP5083/FP4083/FP2083  
Isotropic Field Probe

---

## SPECIFICATIONS

---

|  |   |  |
|--|---|--|
| <b>Dynamic Range:</b>                    | 3 to 300 Volts/meter<br>(V/m)   |  |
| <b>Ranges:</b>                           | 10, 30, 100, 300 V/m<br>full scale  |  |
| <b>Frequency Response:</b>               | 80 MHz to 26 GHz $\pm$<br>1.5 dB  |  |
|  | 26 GHz to 40.0 GHz<br>$\pm$ 3.0 dB  |  |
| <b>Linearity:</b>                        | $\pm$ 0.5 dB full scale<br>(F.S.): $\pm$ 2 least<br>significant bits (LSBs)<br>of A/D converter |  |
| <b>Isotropicity:</b>                     | $\pm$ 1.2 dB  |  |
| <b>Average Overload:</b>                 | 6 times full scale<br>(each range)  |  |
| <b>Environmental:</b>                    | Operating<br>Temperature:   | 10 °C to 40 °C (50 °F<br>to 104 °F)                |
|  | Humidity:   | 5% to 95% relative<br>humidity, non-<br>condensing |
| <b>Fiber Optic Cable<br/>Connectors:</b> | Standard FSMA   |  |
| <b>Battery:</b>                          | 3.6 VDC, 1400 mA-h<br>rechargeable NiCd)  |  |
| <b>Battery Charger:</b>                  | 115/230 VAC,<br>approximately 1 hr  |  |
| <b>Battery Life:</b>                     | 17 Hrs continuous<br>(full charge)  |  |
| <b>Probe Mount:</b>                      | $\frac{1}{4}$ - 20 UNC tapped<br>hole (internal thread)<br>in base of probe                     |  |
| <b>Size:</b>                             | Length (including<br>electronics housing):  | 432 mm (17.0 in)                                   |
|  | Probe diameter:   | 102 mm (4.0 in)                                    |
|  | Weight:   | 0.54 Kg (19 oz.)                                   |
| <b>Optional<br/>Equipment:</b>           | See Table 2   |  |

This page intentionally left blank.

## 2. GETTING STARTED

### UNPACKING AND ACCEPTANCE

Step 1. Upon delivery of your order, inspect the shipping container(s) for evidence of damage. Record any damage on the delivery receipt before signing. In case of concealed damage or loss, retain the packing materials for inspection by the carrier.

Step 2. Remove the probe and battery charger from their shipping containers. Save the boxes and any protective packing materials for future use.

Step 3. Check all materials against the packing list to verify that the equipment received matches what was ordered. If you find any discrepancies, note them and call ETS-Lindgren Customer Service for further instructions.

Be sure that you are satisfied with the contents of your order and the condition of your equipment before installing the probe.


### PROBE

A switch, two fiber optic connectors and a battery charger connector are mounted on the HI-4453/FP5083/FP4083/FP2083 electronics housing (Figure 2).

### **XMIT/RCV**

The fiber optic cable assembly from the receiver is attached to the probe via two connectors. The cables are color-coded—white for XMIT, yellow for RCV. Identically-colored dots are located on the electronics housing adjacent to these connectors. Be sure that each cable is attached to the proper probe connector.

When the fiber optic cables are not attached, always cover the probe connectors with the protective plastic covers supplied with the unit, or with similar material. This prevents dirt or other contaminants from entering the connector and causing communication problems.



**Figure 2: Switch and Connectors**

#### **ARM/OFF**

The ARM/OFF switch activates and deactivates the probe. In the "ARM" position, its internal 3.6 VDC NiCd battery powers the probe: in the "OFF" position, the probe is inactive. To prolong battery life, set the ARM/OFF switch to "OFF" at the end of a test sequence or when the probe is not in use.

#### **Charger**

A standard fast charger is supplied with the HI-4453/FP5083/FP4083/FP2083. When charging is complete, the fast charger acts as a trickle charger. The battery can be left on this maintenance mode indefinitely and its performance will not degrade.

---

Isotropic Electric Field Probes  
HI-4453/FP5083/FP4083/FP2083

### **Battery**

The NiCd battery provides up to 17 hours of probe operation when fully charged.

---

### **BATTERY CHARGING**

Each HI-4453/FP5083/FP4083/FP2083 probe contains a rechargeable nickel-cadmium (NiCd) battery. A fully-charged battery (nominal output voltage of 3.6 VDC) provides up to 17 hours of continuous operation.

**NOTE:** ETS-Lindgren charges the internal NiCd battery of the HI-4453/FP5083/FP4083/FP2083 at the factory in order to calibrate the probe prior to shipment. While every effort is made to ensure that your probe arrives ready to use, we cannot guarantee that this will be the case. Always check the condition of the probe's battery prior to making any measurements.

### **Charging Procedure**

Step 1. Plug the charger into a suitable AC source.

Step 2. Set the probe switch to OFF. Insert the plug on the charger cable into the probe's CHARGER jack.

Step 3. The battery is now charging. This may take approximately one hour, depending on how deeply the batteries are discharged. When charging is complete, the charger automatically goes into a trickle charge and will continue to do so until the probe is disconnected.

### **Battery Tips**

NiCd batteries have several characteristics that can affect both their performance and operating life. The following tips advise you how to take advantage of these characteristics to get the most out of your probe's battery.

- Although NiCd batteries are rated for operation in temperatures from -20°C to +65°C (-4 °F to +140 °F), operating the probe in extreme temperatures will reduce operating time significantly. The optimum operating temperature range for these batteries is +20°C to +30°C (+68°F to +86°F).
- The battery in the HI-4453/FP5083/FP4083/FP2083 probe does not require periodic "deep discharges" to reverse the capacity-depleting "memory effect" caused by repeated shallow discharges; however, undercharging can reduce battery capacity. Therefore, after the charging procedure is complete, be sure that the battery is fully charged before resuming field operation.
- If the battery exhibits low terminal voltage during charging, or if it appears unable to acquire or maintain an appreciable charge, individual cells in the battery may be shorted or damaged. If, for any reason, your battery needs replacement, contact ETS-Lindgren Customer Service for assistance.

### **3. OPERATION**

This section discusses the theory of operation and the functions of the HI-4453/FP5083/FP4083/FP2083 Isotropic Electric Field Probe. A high-level block diagram (Figure 3) is included to aid the discussion. The objective is to provide information enhancing user understanding of the design of this probe.

#### **SYSTEM THEORY**

---

The HI-4453/FP5083/FP4083/FP2083 Broadband Isotropic Electric Field Probe utilizes a microprocessor for intelligent operation and control. The probe's self-contained power supply employs a 3.6 VDC NiCd battery, which provides up to 17 hours of continuous operation.

For each axis, the probe measures the radio frequency signal level and generates a linearized reading of the measurement. The result is transmitted to the readout over glass fiber optic cables. See "Probe Operation" for a more detailed explanation of the functioning of the HI-4453/FP5083/FP4083/FP2083.

The probe relays data to the readout via either a short form or long form output word. See Appendix B for details on both output word formats.

For some applications, you may use a computer with an RS232 serial port to communicate directly with the HI-4453/FP5083/FP4083/FP2083 via ETS-Lindgren's optional HI-4413P Fiber Optic/RS232 Interface.




Figure 3: Operation Block Diagram

## **PROBE OPERATION**

---

Receiver commands to the probe consist of the following:

- Send reading
- Read battery voltage
- Zero
- Set sleep timer
- Switch range
- Read temperature
- Enable axis

The signal flow within the probe is shown in the block diagram.

To measure field strength, three mutually orthogonal dipole antennas (one per axis) are used to provide an isotropic response to the ambient field. The signal from each axis is fed to a Schottky diode detector operating in its Square-Law region. After filtering, the signals generated by each axis are fed to the instrumentation amplifier, whose output feeds the selectable range/offset stage.

For each of the four ranges, the selectable range and offset stage provides a coarse analog zero for the measured signal. The output of the range/offset stage is fed to the A/D multiplexer, then to the analog-to-digital (A/D) converter itself.

After acquiring the composite three-axis signal, the microprocessor commands the A/D multiplexer to read the battery voltage and temperature sensing lines. Data from the A/D converter is fed to the microprocessor, which transmits it to the receiver.

The EEPROM stores all calibration data for the probe.

## **PROBE POWER SUPPLY**

---

A sealed rechargeable 3.6 VDC NiCd battery, which drives both the analog and digital power supplies, powers the HI-4453/FP5083/FP4083/FP2083 probe. Separate power sources provide isolation between the analog and digital circuitry. With the probe switch in the "ARM" position, battery voltage is applied to the power switch, which routes this voltage to the power supply, enabling the microprocessor. A timer circuit controls the power switch. The timer monitors the fiber optic connector input line to determine whether the probe has received a command during a specified period (several seconds). If no command is received during this period, the timer signals the power switch to disable the power supply and the microprocessor. In essence, the probe goes dormant to conserve battery power; only the fiber optic input circuitry remains active in order to detect new commands. When the next command from the receiver reaches the probe, power is reapplied automatically and the processor is reactivated, "waking up" the probe.

**NOTE:** The probe uses volatile random access memory (RAM). If, for any reason, power to the probe is lost, the probe must be re-zeroed.

### **Zeroing**

When the receiver sends a zero command, the probe must be in a zero field environment. This is because the zero command causes the multiplexer (via the processor) to perform a normal read cycle on the composite axis signal. This procedure is executed for all ranges. When the processor receives all the zero-field values, it stores them in a special register; these values are subtracted from all subsequent measurements. Therefore, a probe that is zeroed while it is not in a zero field environment will give erroneous readings.

## **APPLICATION CONSIDERATIONS**

The following subsections contain information designed to help you maximize the effectiveness of the HI-4453/FP5083/FP4083/FP2083.

### **Out-of-Band Considerations**

Although the HI-4453/FP5083/FP4083/FP2083 is nominally rated for operation from 80 MHz to 40 GHz, it responds to signals both above and below these frequencies. Such responses must be taken into account when performing certain operations, such as zeroing.

On the lower end, the HI-4453/FP5083/FP4083/FP2083 can exhibit significant response to frequencies below 80 MHz. Such out-of-band responses pose a problem when zeroing the unit, since that operation obviously assumes a zero field condition. At the upper end, similar problems can occur. The specified upper operating limit for this probe is 40 GHz. Above this limit there will be some response, though it will not be consistent or predictable. Any probe response to out-of-band frequencies must be compensated for when zeroing the probe.

### **Resolution Limitations**

Limitations in probe resolution may result in a non-zero reading when the receiver is zeroed. If this occurs, it does not necessarily mean either that there is a problem with the receiver or that your readings are inaccurate. Receiver linearity is specified as + 0.5 dB full scale; in addition, the variance of the A/D converter is + 2 least significant bits. When using the most sensitive range on a given probe, these specifications create the possibility that, under zero field conditions, the receiver may display a non-zero value.

### **Probe Support Structures**

It is very important to keep conductive objects away from the HI-4453/FP5083/FP4083/FP2083. Any such objects in the proximity of the probe may distort the near field and compromise measurement accuracy. If your application requires measurements from a fixed position, always mount the probe on a non-metallic platform, using non-metallic screws.

## 4. MAINTENANCE

This section explains which maintenance tasks can be performed by the user. It also provides information regarding replacement and optional parts. If you have any questions concerning probe maintenance, consult ETS-Lindgren Customer Service.

### **MAINTENANCE RECOMMENDATIONS**

Maintenance of the HI-4453/FP5083/FP4083/FP2083 probe is limited to external components such as cables or connectors.

Any calibration or maintenance task that requires probe disassembly should be performed at the factory. Check with ETS-Lindgren Customer Service (+1-512-531-6400) before opening the unit to avoid problems with your probe's warranty.

**NOTE:** Opening the probe enclosure may void your warranty. If your system is still under warranty, contact ETS-Lindgren Customer Service before performing any maintenance inside the probe.

### **UPGRADE POLICIES**

Periodically, probes are upgraded to enhance functionality. These upgrades are commonly announced through ETS-Lindgren Engineering Bulletins.

### **RETURN PROCEDURES**

To return a system or system component, use the following procedure.

**Step 1.** Contact ETS-Lindgren Customer Service (+1-512-531-6400) to obtain an SRO, Service Request Order.

**Step 2.** Briefly describe the problem in writing. Give details regarding the observed symptom(s) or error codes, and whether the problem is constant or intermittent in nature. Please include the date(s), the service representative you spoke with, and the nature of the conversation. Include the serial number of the item being returned.

**Step 3.** Package the system or component carefully. Use the original packing boxes and materials, if possible.

If the probe is under warranty, refer to the Limited Warranty section of this manual for additional information about your return.

### **PERIODIC/PREVENTIVE MAINTENANCE**

The HI-4453/FP5083/FP4083/FP2083 probe assembly (sensor and electronics housing) requires an annual calibration to verify that they are performing within specifications. ETS-Lindgren service personnel may perform this calibration at the factory. Return your probe(s), using the original packing materials (if possible), to:

**ETS-Lindgren**  
Attn. Service Department  
1301 Arrow Point Drive  
Cedar Park, Texas, USA 78613

## PARTS INFORMATION

---

Use the following tables for ordering replacement (Table 1) or optional (Table 2) parts for HI-4453/FP5083/FP4083/FP2083 probes.

| Part Description<br>(Replacement Parts)                 | Part Number |
|---|-------------|
| Battery Pack, 3.6 VDC,<br>Rechargeable                  | H-491038    |
| Standard Fast Charger<br>(115/230 Volt)                 | H-491198-36 |
| Cable, Fiber Optic, Glass,<br>where XX=length in meters | H-490994-xx |
| Probe Cover (3 required)                                | H-470286A   |
| Handle Assembly   | H-491073    |
| HI-4453/FP5083/FP4083/<br>FP2083 User's Manual          | H-600064    |

Table 1: Replacement Parts List

| Part Description (Optional<br>Parts)                 | Part Number |
|--|-------------|
| Probe Support, HI-<br>4453/FP5083/FP4083/FP2083      | H-490984    |
| Tripod, Dielectric, HI-4453/<br>FP5083/FP4083/FP2083 | H-491000    |
| Fiber Optic / RS232 Interface                        | HI-4413P    |
| Probe Stand  | H-491269    |
| Graphical EMF Readout                                | HI-4460     |
| Numeric EMF Readout                                  | HI-4416     |
| Field Monitor  | FM5004      |

Table 2: Optional Parts List

This page intentionally left blank.

## **5. WARRANTY POLICY FOR ELECTROMAGNETIC ISOTROPIC FIELD MONITORING COMPONENTS**

### **SCOPE AND DURATION OF WARRANTIES**

---

Seller warrants to Buyer that the Holaday Brand Products to be delivered hereunder will be (1) free from defects in material, manufacturing workmanship, and title, and (2) conform to the Seller's applicable product descriptions and specifications, if any, contained in or attached to Seller's quotation. If no product descriptions or specifications are contained in or attached to the quotation, Seller's applicable product descriptions and specifications in effect on the date of shipment shall apply. The criteria for all testing shall be Seller's applicable product specifications utilizing factory-specified calibration and test procedures and instruments.

All product warranties, except the warranty of title, and all remedies for warranty failures are limited in time as shown in the table below.

| <u>Product Warranted</u> | <u>Duration of Warranty Period</u> |
|--------------------------|------------------------------------|
| FM5004                   | 3 Years                            |
| HI-6005                  | 3 Years                            |
| HI-6001                  | 3 Years                            |
| HI-4450                  | 3 Years                            |
| FP-4080                  | 3 Years                            |
| FP-5000                  | 3 Years                            |
| FP-2080                  | 3 Years                            |
| HI-4453                  | 3 Years                            |
| FP-4083                  | 3 Years                            |
| FP-5083                  | 3 Years                            |
| FP-2083                  | 3 Years                            |
| FP-5080                  | 3 Years                            |
| FP-4000                  | 3 Years                            |
| FP-2000                  | 3 Years                            |
| HI-4422                  | 3 Years                            |

Any product or part furnished to Buyer during the warranty period to correct a warranty failure shall be warranted to the extent of the unexpired term of the warranty applicable to the repaired or replaced product.

The warranty period shall commence on the date the product is delivered to Buyer; however, if Seller assembles the product, or provides technical direction of such assembly, the warranty period for such product shall commence on the date the assembly of the product is complete. Notwithstanding the foregoing, in the event that the assembly is delayed for a total of thirty (30) days or more from the date of delivery for any reason or reasons for which Seller is not responsible, the warranty period for such product may, at Seller's options, commence on the thirtieth (30th) day from the date such product is delivered to Buyer. Buyer shall promptly inspect all products upon delivery. No claims for shortages will be allowed unless shortages are reported to Seller in writing within ten (10) days after delivery. No other claims against Seller will be allowed unless asserted in writing within thirty (30) days after delivery (or assembly if the products are to be assembled by Seller) or, in the case of alleged breach of warranty, within the applicable warranty period.

### **WARRANTY EXCLUSIONS**

Except as set forth in any applicable patent indemnity, the foregoing warranties are exclusive and in lieu of all other warranties, whether written, oral, express, implied, or statutory. EXCEPT AS EXPRESSLY STATED ABOVE, SELLER MAKES NO WARRANTY, EXPRESS OR IMPLIED, BY STATUTE OR OTHERWISE, WHETHER OF MERCHANTABILITY OR FITNESS FOR ANY PARTICULAR PURPOSE OR USE OR OTHERWISE ON THE PRODUCTS, OR ON ANY PARTS OR LABOR FURNISHED DURING THE SALE, DELIVERY OR SERVICING OF THE PRODUCTS. THERE ARE NO WARRANTIES WHICH EXTEND BEYOND THE DESCRIPTION ON THE FACE HEREOF.

Warranty coverage does not include any defect or performance deficiency (including failure to conform to product descriptions or specifications) which results, in whole or in part, from (1) negligent storage or handling of the product by Buyer, its employees, agents, or contractors, (2) failure of Buyer to prepare the site or provide an operating environmental condition in compliance with any applicable instructions or recommendations of Seller, (3) absence of any product, component, or accessory recommended by Seller but omitted at Buyer's direction, (4) any

---

Isotropic Electric Field Probes  
HI-4453/FP5083/FP4083/FP2083

design, specification, or instruction furnished by Buyer, its employees, agents or contractors, (5) any alteration of the product by persons other than Seller, (6) combining Seller's product with any product furnished by others, (7) combining incompatible products of Seller, (8) interference with the radio frequency fields due to conditions or causes outside the product as furnished by Seller, (9) improper or extraordinary use of the product, or failure to comply with any applicable instructions or recommendations of Seller, or (10) acts of God, acts of civil or military authority, fires, floods, strikes or other labor disturbances, war, riot, or any other causes beyond the reasonable control of Seller. This warranty does not cover batteries or any item that is designed to be consumable. Seller does not warranty products of others, which are not included in Seller's published price lists.

---

### **BUYER'S REMEDIES**

---

If Seller determines that any product fails to meet any warranty during the applicable warranty period, Seller shall correct any such failure by either, at its option, repairing, adjusting, or replacing without charge to Buyer any defective or nonconforming product, or part or parts of the product. Seller shall have the option to furnish either new or exchange replacement parts or assemblies.

Warranty service during the applicable warranty period will be performed without charge to Buyer within the contiguous 48 United States during Seller's normal business hours. After the warranty period, service will be performed at Seller's prevailing service rates. Subject to the availability of personnel, after-hours service is available upon request at an additional charge. For service outside the contiguous 48 United States, travel and per diem expenses, when required, shall be the responsibility of the Buyer, or End User, whichever is applicable.

The remedies set forth herein are conditioned upon Buyer promptly notifying Seller within the applicable warranty period of any defect or nonconformance and making the product available for correction.

The preceding paragraphs set forth Buyer's exclusive remedies and Seller's sole liability for claims based on failure of the products to meet any warranty, whether the claim is in contract, warranty, tort (including negligence and strict liability) or otherwise, and however instituted, and, upon the expiration of the applicable warranty period, all such liability shall terminate. IN NO EVENT SHALL SELLER BE LIABLE TO BUYER FOR ANY SPECIAL INDIRECT, INCIDENTAL OR CONSEQUENTIAL DAMAGES OF ANY KIND ARISING OUT OF, OR AS A RESULT OF, THE SALE, DELIVERY, NON-DELIVERY, SERVICING, ASSEMBLING, USE OR LOSS OF USE OF THE PRODUCTS OR ANY PART THEREOF, OR FOR ANY CHARGES OR EXPENSES OF ANY NATURE INCURRED WITHOUT SELLER'S WRITTEN CONSENT DESPITE ANY NEGLIGENCE ON BEHALF OF THE SELLER. IN NO EVENT SHALL SELLER'S LIABILITIES UNDER ANY CLAIM MADE BY BUYER EXCEED THE PURCHASE PRICE OF THE PRODUCT IN RESPECT OF WHICH DAMAGES ARE CLAIMED. This agreement shall be construed in accordance with laws of the State of Illinois. In the event that any provision hereof shall violate any applicable statute, ordinance, or rule of law, such provision shall be ineffective to the extent of such violation without invalidating any other provision hereof.

Any controversy or claim arising out of or relating to the sale, delivery, nondelivery, servicing, assembling, use or loss of use of the products or any part thereof or for any charges or expenses in connection therewith shall be settled in Austin, Texas by arbitration in accordance with the Rules of the American Arbitration Association, and judgment upon the award rendered by the Arbitrator may be entered in either the Federal District Court for the Western District of Texas or the State District Court in Austin, Texas, all of the parties hereto consenting to personal jurisdiction of the venue of such court and hereby waive the right to demand a jury trial under any of these actions

## 6. APPENDIX A: ERROR CODES

### **PROBE ERROR OUTPUT**

---

If an error occurs, the probe will respond with one of the following strings. These strings begin with a colon and end with a carriage return.

|     |  |
|-----|--|
| E01 | Communication error (e.g., overflow).  |
| E02 | Buffer full error. Too many characters contained between the Start Character/Carriage Return sequence. |
| E03 | The received command is not valid.   |
| E04 | The received parameter is not valid.   |
| E05 | Hardware error (e.g., EEPROM failure).   |
| E06 | Parity error.  |

This page intentionally left blank.

## **7. APPENDIX B: HI-4453/FP5083/FP4083/FP2083 OPERATING PROTOCOLS**

The information in this appendix assumes that you have purchased the optional HI-4413P Fiber Optic/RS232 Interface, and are capable of communicating directly with the HI-4453/FP5083/FP4083/FP2083 probe via computer. No system readout is required when using this configuration.

### **COMMUNICATION PROTOCOL**

---

|                     |               |
|---------------------|---------------|
| <b>Data Type:</b>   | RS-232 Serial |
| <b>Data Mode:</b>   | Asynchronous  |
| <b>Word Length:</b> | 7 bit         |
| <b>Parity:</b>      | Odd           |
| <b>Stop Bits:</b>   | 1             |
| <b>Data Rate:</b>   | 9600 baud     |

### **INFORMATION TRANSFER PROTOCOL**

---

The HI-4453/FP5083/FP4083/FP2083 operates as a Controller Mode device. This probe only responds to commands from another device; it transmits no data without receiving instructions to do so.

#### **Command Structure**

A command to an HI-4453/FP5083/FP4083/FP2083 probe consists of

- 1) A command letter, followed by
- 2) Possible parameters
- 3) Terminated with a carriage return.

When it completes the command, the probe responds with a string consisting of

- 1) A start character ":")
- 2) The command letter, followed by
- 3) Data, if required, and terminated with
- 4) A carriage return.

If the command does not require the probe to return any data, the probe simply responds with the command letter and a carriage return. If an error occurs, the probe responds with an error code, as detailed in Appendix A.

## **COMMANDS**

---

| <b>Command</b> | <b>Description</b>  |
|----------------|---|
| Axxx           | Axis enable/disable. x = "E" means enable, x = "D" means disable. xxx order is X axis, Y axis, Z axis   |
| B              | Read battery voltage.   |
| Cx             | Set baud rate.<br>x = 1 sets rate to 2400 baud<br>x = 2 sets rate to 9600 baud<br>New baud rate does not take effect until the next power-up. |
| Dx             | Read probe data.<br>x = 1 enables short form output<br>x = 2 enables long form output   |
| Lx             | Load table data.<br>x = ASCII hex data  |
| Rx             | Set range.<br>x = 1, 2, 3, 4 or N (next range)  |
| Sx             | Sleep timer.<br>x = number of seconds to wait for a command before putting the probe into the sleep mode.                                     |
| Tx             | Read Temperature.<br>x = C or F   |
| Ux             | Set unit type.<br>x = 1, 2, 3, or N (next unit)<br>1 = V/m<br>2 = mW/cm <sup>2</sup><br>3 = [V/m] <sup>2</sup>                                |

---

 Isotropic Electric Field Probes  
 HI-4453/FP5083/FP4083/FP2083

| Command | Description  |
|---------|--|
| Vtx     | Verify table data.<br>t = table (same table as for the Lx command, above). x = number of bytes to send in one string.  |
| Z       | Zero.  |
| Null    | Send the ASCII null character. This is a special command that can be used as the initial command to the probe after it is powered up. The probe responds with "N". |

**Probe Output**

| Command | HI-4453/FP5083/FP4083/FP2083 Response   |
|---------|---|
| B       | Bxx.xx, where xx.xx is the battery voltage.   |
| D1      | Dxx.xxuuu, the short form output.<br>xx.xx is the reading. The position of the decimal point depends upon the range setting of the probe.   |
|         | uuu = units<br>$\_V\_{}$ = V/m, mW2 = mW/cm <sup>2</sup> , $\_V^2\_{}$ = [V/m] <sup>2</sup> (underscore indicates a space character)  |
| D2      | Dxx.xxuuurrrobaat, the long form output.<br>xx.xx = the reading, as described for D1.<br>uuu = units, as describe for D1.<br>rrr = recorder out value (A 3-digit ASCII number from 0 to 255).<br>o = over range indicator ("N" = ok, "O" = over range). |
|         | b = battery status ("N" = safe operating level, "W" = warning level, "F" = fail level).<br>aaa = axis enable ("E" = enabled).<br>t = terminating carriage return.   |
| Rx      | Rx, where x is the range.<br>x = "" returns the range currently in use<br>x = 1, 2, 3, 4 enables the selected range<br>x = N sets the probe to the next (higher) range.   |
| TF      | Txxx, where xxx is temperature in ° Fahrenheit.   |
| TC      | Txxx, where xxx is temperature in ° Centigrade.   |
| Vtx     | Vxx[xx...], where xx is table data in ASCII hex.  |

This page intentionally left blank.

**8. APPENDIX C: SERIES 491198-36  
BATTERY FAST CHARGER  
MANUAL**

This page intentionally left blank.

# Series 491198-36

## NiCd Battery Fast Charger

### User's Manual

#### Declaration of Conformi



ETS-Lindgren, L.P.  
1301 Arrow Point Drive  
Cedar Park, TX 78613  
USA

Declares that the HOLADAY product described in this instruction manual is in compliance with: EN EMC Directive 86/336/EEC, EN50082-1, EN55011

**ETS-Lindgren, L.P.**  
(Formerly Holaday Industries, Inc.)



ETS-Lindgren L.P. reserves the right to make changes to any product described herein in order to improve function, design or for any other reason. Nothing contained herein shall constitute ETS-Lindgren L.P. assuming any liability whatsoever arising out of the application or use of any product or circuit described herein. ETS-Lindgren L.P. does not convey any license under its patent rights or the rights of others.

© Copyright 2004 by ETS-Lindgren L.P. All Rights Reserved.  
No part of this document may be copied by any means  
without written permission from ETS-Lindgren L.P.

**EMAIL & INTERNET**

support@ets-lindgren.com  
<http://www.ets-lindgren.com>

**USA**

1301 Arrow Point Dr., Cedar Park TX 78613 USA  
P.O. Box 80589, Austin, TX 78708-0589 USA  
Tel 512.531.6400  
Fax 512.531.6500

**FINLAND**

Mekaanikontie 1, 27510, Eura, Finland  
Tel 358.2.838.3300  
Fax 358.2.865.1233

**JAPAN**

4-2-6, Kohinata  
Bunkyo-ku  
Tokyo 112-0006 JAPAN  
Phone + 81 3 3813 7100  
Fax + 81 3 3813 8068

**CHINA**

1917-1918 Xue Zhixuan Building  
No 16 Xue Qing Road  
Haidian District  
Beijing Postcode: 100083 CHINA  
Phone + 86 10 82755304  
Fax + 86 10 82755307

---

Isotropic Electric Field Probes  
HI-4453/FP5083/FP4083/FP2083

Table of Contents

|     |                                     |    |
|-----|-------------------------------------|----|
| 1.0 | General Safety Considerations ..... | 41 |
| 2.0 | Description.....                    | 43 |
| 3.0 | UNPACKING AND ACCEPTANCE .....      | 45 |
| 4.0 | Specifications.....                 | 47 |
| 5.0 | OPERATING INSTRUCTIONS .....        | 51 |
| 6.0 | MAINTENANCE .....                   | 55 |

---

Isotropic Electric Field Probes  
HI-4453/FP5083/FP4083/FP2083

Revision Record  
Manual #600086

Series 491198-36 NiCd Battery Fast Charger

| <u>Revision</u> | <u>Description</u>           | <u>Date</u>   |
|-----------------|------------------------------|---------------|
| A               | Release                      | 4/99          |
| B               | Changed Area Code<br>Updates | 2/00<br>12/04 |

## 1.0 General Safety Considerations



This symbol alerts the user that important literature concerning the operation and maintenance of this unit has been included. Therefore, it should be read carefully in order to avoid any problems.

Warning

This is a Safety Class I product (provided with a protective earthing ground incorporated in the power cord). The mains plug shall only be inserted in a socket outlet provided with a protective earth contact. Any interruption of the protective conductor, inside or outside the instrument, is likely to make the instrument dangerous. Intentional interruption is prohibited. DO NOT defeat the earth-grounding protection by using an extension cable, power cable, or autotransformer without a protective ground conductor

This instrument is to be used with a three-wire power cord set which meets or exceeds the requirements of EN60799. The power cord set used must be rated for a minimum of 250V/10A. When connected to an appropriate power line outlet, this cable grounds the instrument cabinet

Warning

No operator serviceable parts exist inside the device. Refer servicing to qualified personnel. To prevent electrical shock, do not remove cover.

Warning

If this instrument is used in a manner not specified by ETS-Lindgren, the protection provided by the instrument may be impaired.

Warning

This battery charger incorporates parts, such as a switch and relay that potentially could produce sparks or arcs.

Warning

For indoor use only, do not expose to rain.




This page intentionally left blank.

## 2.0 Description

The 491198-36 NiCd Battery Fast Charger is a dual power source battery charger. It charges 3.6 volt 1400 mAH NiCd batteries and is powered by 120-240 Vac line power or 12.5 Vdc. The 491198-36 uses a  $-(dV)/(dT)$  (negative delta V) technique to determine when the battery is fully charged, typically one hour. With this technique, the charge state of the battery has no effect other than shortening the charge time.

Housed in a rugged enclosure, power enters the battery charger through a power entry module, which contains the fuses, or an optional cigarette lighter plug adapter. LEDs and the label on the front face of the unit provide operating status. The battery charger connects to the device being charged through a short cord terminated with a power jack.

An integrated circuit within the charger monitors the battery voltage and controls the charging functions according to the charge state of the battery.



**Figure 1: 3.6V Battery Charger**

This page intentionally left blank.

## 3.0 UNPACKING AND ACCEPTANCE

### **INTRODUCTION**

---

This section contains information on unpacking and acceptance of the 491198-36 NiCd Battery Fast Charger.

Step 1. Upon delivery of your order, inspect the shipping container(s) for evidence of damage. Record any damage on the delivery receipt before signing. In case of concealed damage or loss, retain the packing materials for inspection by the carrier.

Step 2. Remove the product from its shipping container. Save boxes and any protective packing materials for future use.

Step 3. Check all materials against the packing list to verify that the equipment received matches that which was ordered. If you find any discrepancies, note them and call ETS-Lindgren Customer Service for further instructions.

Be sure that you are satisfied with the contents of your order and the condition of your equipment before using it.



## 4.0 Specifications

### **POWER**

---

Main: IEC filtered AC power input module  
110 - 240 Vac, 500 mA MAX., 50 - 100 Hz  
Alternate: Automobile cigarette lighter to 2mm power plug  
adapter cord., 12.5 Vdc, 100 mA  
Fuses: 250 Volt, 1.0 Amp, Type T (5 mm x 20 mm)

### **OUTPUT**

---

Open Circuit Voltage: 15 Vdc  
Fast Charge Pending Current: 60 mA  
Fast Charge Current: 1400 mA  
Pulsed Trickle Charge Current: 50 mA  
Output Voltage (During Fast Charge): 3 - 6 Vdc

### **NICD BATTERY:**

---

3.6 Volt 3 Cell NiCd Battery, 1400 mAH (Rapid charge cells, 1.2  
volts/cell); ETS-Lindgren Part #491038

### **ENVIRONMENTAL**

---

Operating temperature 10 to 40° C  
Humidity: 5% to 95% relative humidity,  
non-condensing

### **POWER CABLE**

---

This charger is shipped with a three-wire power cable. When  
this cable is connected to an appropriate AC power source, it  
connects the chassis to earth ground. The type of power cable  
shipped with each battery charger depends on the country of  
destination.

## POWER CABLE SET INFORMATION

| ETS-<br>Lindgren<br>Part No. | Manufacturer, Part No.<br>and Information   |
|------------------------------|---|
| 2217500                      | Volex #17500<br>Type SVT, Foil shielded, PVC Jacketed, 60°C<br>Molded PVC Grounding Plug NEMA 5-15P UC-004<br>Molded PVC Receptacle IEC320/C13 UC-005<br>18 Ga. 3 Cond. 10A-125V  |
| 2217506C                     | Volex #17506<br>Type SVT, Foil shielded, PVC Jacketed, 60°C<br>Molded PVC Grounding Plug NEMA 5-15P UC-004<br>Molded PVC Right-Angle Receptacle<br>IEC320/C13 UC-006<br>18 Ga. 3 Cond. 10A-125V<br>Countries: Canada, Japan, Puerto Rico, Taiwan, Venezuela, Hong Kong, United States   |
| 221100                       | Kobiconn #173-7001<br>Type H05VV-F, PVC Jacketed, 70°C<br>Molded PVC Right-Angle Grounding Plug CEE 7/7 UC-814<br>Molded PVC Receptacle IEC320/C13 UC-051<br>1.0mm <sup>2</sup> 3 Cond. 10A-250V<br>Countries: Argentina, Austria, Brazil, Finland, France, Germany, Israel, Italy, Holland, Korea, Netherlands, Norway, Sweden, Turkey |
| 222600                       | QualTek #370001-E01<br>Type H05VV-F, PVC Jacketed, 70°C, Harmonized<br>Molded PVC Grounding Plug BS 1363, Fused UC-851<br>Molded PVC Receptacle IEC320/C13 UC-852<br>1.0mm <sup>2</sup> 3 Cond. 10A-250V<br>Countries: England, Ireland, Malaysia, Scotland, Singapore, South Africa, Wales   |
| 221600                       | Leeds Electronic Components #FFBS-1310<br>Type SAA, Ordinary Duty, PVC Jacketed, 75°C<br>Molded Grounding Plug AS3112 UC-822<br>Molded PVC Receptacle IEC320/C13 UC-051<br>1.0mm <sup>2</sup> 3 Cond. 10A-250V  |

---

Isotropic Electric Field Probes  
HI-4453/FP5083/FP4083/FP2083

| ETS-<br>Lindgren<br>Part No. | Manufacturer, Part No.<br>and Information  |
|------------------------------|--|
| 221500                       | Countries: Australia, China<br>Volex #2102H-C3-10<br>Type H05VV-F, PVC Jacketed, 70°C<br>Molded PVC Grounding Plug SEV 1011 UC-841<br>Molded PVC Receptacle IEC320/C13 UC-051<br>1.0mm <sup>2</sup> 3 Cond. 10A-250V<br>Country: Switzerland |

This page intentionally left blank.

## 5.0 OPERATING INSTRUCTIONS

The 491198-36 battery charger is intended to charge the 3.6 volt NiCd batteries, either in the lab or in the field.

### **INPUT POWER REQUIREMENTS**

The 491198-36 charger may be powered by standard line voltage (110 - 240 Vac, 50 - 60 Hz) or by an optional automobile cigarette lighter plug (12.5 Vdc), see specifications. The AC power entry module contains a fuse.

### **FUSE REPLACEMENT**


If a fuse has blown, it must be replaced with the same type and value or an unsafe condition may result. Refer to Figure 2 for fuse replacement.

The fuse is contained in the fuse drawer in the power input module. To access the fuse, use a screwdriver to pry the drawer open and remove it from the module. The drawer holds two fuses; the fuse towards the outside of the drawer is a spare.

After the fuse has been replaced, slide the fuse drawer back into the module. Make sure that it snaps securely into its locked position.

**CAUTION**

NEVER attempt to change the fuse with the battery charger plugged in.



**Figure 2: Fuse Replacement**

## **CHARGING A BATTERY**

After connecting the battery charger to a proper power source, the battery charger simply plugs into the charger jack on the HI-6005. The HI-6005 must be turned off or the battery will not charge.

### **CAUTION**

NEVER attempt to recharge a non-rechargeable battery.

## **CHARGING INDICATORS**

---

There are five LEDs located on the front of the charger that provide operating information to the user.

The “POWER ON” LED (green) will remain illuminated as long as the charger is plugged into the AC power source.

If the charger does not detect a battery, the “NO BATTERY” LED (amber) light will illuminate.

When the charger does detect the unit’s battery, the “PENDING” LED (amber) light illuminates while the charger qualifies the battery for fast charge. If the voltage is below the safe fast charge level, the battery is charged in the pulse trickle mode.

When the voltage is at a safe level the charger will switch to the fast charge mode and the “CHARGING” LED (amber) light illuminates.

When charging is complete, the charger switches back to the pulse trickle mode and the “COMPLETE” LED (green) light will illuminate. The device can be left on the charger in this maintenance mode indefinitely.



## 6.0 MAINTENANCE

Electronic instruments are delicate. Operate the battery charger with care.

User serviceable parts do not exist inside the battery charger. Warranty may be void if the battery charger housing is opened.

If the battery charger fails to operate, check for a blown fuse inside the power entry module (refer to figure 2).

**CAUTION**

NEVER attempt to change the fuse with the battery charger plugged in.

If a fuse is blown it must be replaced. Be sure to use only 250 Volt, 1.0 Amp, Type T (5 mm x 20 mm).

If the battery charger still fails to operate, or if you have any questions concerning charging your products, contact ETS-Lindgren Customer Service.